



Article History

Submitted: 28-02-2026

Revised: 08-04-2026

Accepted: 13-04-2026

Corresponding Author

Salman Kunnipurayil

Salman.vcfg31@gmail.com

Quality Control and Process Validation in Medical Device Manufacturing

Salman Kunnipurayil^{1*}, Mannan Mohammed²

¹Northern Illinois University

²St Mary's University San Antonio, TX

¹Salman.vcfg31@gmail.com, ²mohammedm9512@gmail.com

Abstract

Medical device manufacturers must incorporate quality control and process validation as a part of the process, which guarantees the safety of the product, its reliability, and regulatory compliance. Quality control aims at identifying defects by inspection and testing and process validation comes up with documented evidence that the manufacturing processes continuously produce devices within specifications that have been defined. Collectively, they create a system of quality that is integrated and reduces the risks and improves the consistency of products. This review identifies regulatory frameworks, methodologies, tools, challenges, and new trends of these practices. As the complexity of devices increases, and global regulations become stricter, implementing new technologies and lifecycle-oriented approach is essential to continue upholding high-quality standards and enhancing patient safety outcomes.

Key words

Quality control, process validation, medical devices, regulatory compliance, FDA guidelines, manufacturing processes, quality assurance.

Introduction

The medical device production sector is one of the most important in the contemporary healthcare system as it supplies the means and technologies involved in diagnosis, treatment, monitoring and rehabilitation. Some of these are as simple as thermometers and syringes, others are as complex as implantable pacemaker devices, imaging devices and robotic surgical instruments [1]. Since they have a direct impact on patient safety and clinical outcomes, the continued quality and reliability of medical equipment is essential to ensure. In this case, quality control and process validation is a crucial element of the manufacturing cycle [2].

Quality control (QC) may be described as the techniques and practices which are put in place to ensure that the products that are manufactured are up to the pre-determined specifications and



regulatory standards. It involves inspection, testing and checking of products in various phases of production to detect defect, deviation or inconsistency [3]. In medical equipment, QC does not only pertain to the inspection of the final product, but also the whole manufacturing process including evaluation of raw materials, in-process inspection and post-production inspection. A good quality control is useful in reducing risk, product recalls, and in ensuring that the devices will perform as intended in the specified conditions [4].

Process validation on the other hand is documented such that, it is used as indications that a manufacturing process has not changed in its ability to manufacture products of a certain quality. As opposed to quality control, which is concerned with identifying faults, process validation is concerned with avoiding them by creating a high degree of confidence with the process itself [5]. It involves a series of activities that involve installation qualification (IQ), operational qualification (OQ), and performance qualification (PQ) that are all geared towards testing the different characteristics of the manufacturing system [6]. It is especially critical to validate the process of a medical device because, in most instances, e.g. sterilization, molding or coating, cannot be fully verified through any end-product testing.

The sophistication of the medical equipment has been on the rise, as well as the high standards of the regulations, so the role of the effective quality systems has been increasing. The quality conditions of product management include the demands of regulatory bodies worldwide that guarantee safety and efficacy of products [7]. The manufacturers, in turn, have to integrate quality management, process validation into a unified framework that could be used to facilitate the process of continuous improvement and risk management.

This review article aims to explore the concepts, theories and significance of quality control and process validation in the production of medical devices. It highlights the synergy of these two elements to ensure quality of products, regulations and patient safety in addition to briefly touching on the modern challenges and new trends in the industry.



Overview of Medical Device Manufacturing

The production of medical equipment is a very regulated and specialized sector and involves designing, developing, manufacturing and supplying of devices within a medical context. These products are of a wide range, with low risk and simple (e.g. bandages, surgical gloves) and high-risk (e.g. implantable defibrillator, diagnostic imaging, minimally invasive surgical devices) products [8]. The fact that this sector has a wide variety of products necessitates manufacturers to have flexible and yet powerful production systems depending on the type of devices and their intended applications.

The manufacturing process normally begins with the product design and development whereby the engineers and scientists translate clinical requirements to the specifications of a working device. The phase entails the selection of materials, prototyping and checking of the design to ensure that the device works as required [9]. Once the design is finished the process is redirected to production planning phase where manufacturing processes, equipment and quality are developed. This step is necessary since the quality of products, their scalability, and their regulation-adherence have a direct effect on the initial decisions [10].

Production is also a process that involves several stages i.e., the sourcing of the raw materials, manufacturing of the parts, assembly, packaging and labeling. Depending on the type of device, manufacturing may be injection molding, extrusion, machining, electronic assembly and sterilization. Monitoring of all these processes must be done to bring uniformity as well as quality assurance [11]. Indicatively, medical equipment should be sterile and this requires proven processes of sterilizing equipment e.g. using ethylene oxide, or radiations to ensure that the equipment has no microorganism that can cause harm to the body [12]. A significant characteristic of manufacturing of medical equipment is that it adheres to the quality management systems and standards.

The specifications of a full quality management system which is peculiar to medical devices are given in ISO standards such as ISO 13485. The regulatory bodies such as the U.S. Food and Drug Administration (FDA) and the European Medicines Agency (EMA) have elaborate guidelines that



the manufacturers have to adhere to during the lifecycle of the product. These laws ensure that machines are safe, efficacious and manufactured under controlled environments [13].

Risk management is the other aspect of the manufacturing process. Manufacturers need to determine the possible risks of using the devices and put measures in place to curb such risks. This includes continuous monitoring, documenting and optimization of the processes to ensure that the products are consistent [14]. It also comes with traceability systems to track materials and components in the supply chain to enable an easy recall or corrective action, should it be needed.

The development of technology has had a major impact on the production of medical devices. The use of robotics, automation, computerized monitoring and tracking systems has increased the efficiency of production and reduced human error. New technologies such as additive manufacturing (3D printing) that enable fast prototyping and customization of devices are also changing the industry [14]. Medical devices are complex and multidisciplinary to manufacture and must be highly coordinated in its design, manufacturing, quality assurance and compliance. It has been ultimate in their efforts to offer safe, reliable and high quality equipment that meets the needs of patients and health professionals [15].

Regulatory Framework and Standards

The medical device market is highly regulated, with the regulatory environment being structured in such a way that it assures all products are safe, effective and of high quality, before they reach the patients. Regulatory frameworks and standards offer a collection of rules, guidelines, and best practices that manufacturers need to observe during the entire lifecycle of the product, which includes design and development to production, distribution, and post-market surveillance of the product [16]. Such rules are necessary since medical equipment has a direct influence on the human health and any defect or failure may have critical results.

International Organization for Standardization standard ISO 13485 is also one of the most well-known international standards in producing medical devices. This standard describes the quality management system (QMS) requirements to a quality management system that is specifically designed to meet the needs of the medical device industry [17]. It focuses on risk management,



process control, documentation and continuous improvement. Conformity to ISO 13485 indicates that a manufacturer has put into place a systematic method of quality and regulatory compliance on products and it is commonly a requirement of entering international markets [18].

The Food and Drug Administration (FDA) is a key regulatory organ in the United States in regulating medical devices in the Federal Food, Drug, and Cosmetic Act. FDA groups medical devices into three different classes (Class I, II, and III) according to the degree of risk they present to the patient. The regulatory requirements imposed on each type of class vary, with general controls on low-risk devices and premarket approval (PMA) on high-risk devices [19]. FDA also implements the Quality System Regulation (QSR), which includes current good manufacturing practices (cGMP) that manufacturers should adhere to, to maintain the same quality of products.

Within Europe, medical devices fall under the Medical Device Regulation (MDR), which supersedes the previous Medical Device Directive (MDD). The MDR is more demanding in terms of clinical evaluation, post-market surveillance, and traceability. Before marketing their products in the European Economic Area, manufacturers are required to acquire the CE marking to ensure that they meet the European safety and performance standards [20]. These requirements are evaluated and certified by the use of notified bodies, which are independent bodies appointed by regulatory bodies.

Outside of the U.S. and Europe, most countries have developed their own set of regulations which, in many cases, follow the international standards to enable international business to occur. As an illustration, the regulatory frameworks of such countries as Canada, Japan, and Australia include the aspects of ISO standards and the risk-based classification systems [21]. Global harmonization initiatives (including those of the International Medical Device Regulators Forum (IMDRF)) are aimed at minimizing regulatory differences and facilitating uniform safety standards across the world [22].




Figure 1. Global Medical Device Regulatory Ecosystem and Lifecycle Integration

Documentation and traceability is an important aspect of regulatory compliance. Manufacturers must keep an elaborate account of design controls, risk assessment, validation activities and quality control processes [23]. Such documentation will provide transparency and will allow regulatory bodies to audit operations and check compliance. Also, post-market surveillance system is applied in order to track the performance of the device once it is commercialized so that manufacturers can quickly discover and correct possible problems [24]. Regulatory frameworks and standards have been the foundation of medical device production in that they set expectations of safety, quality and performance. Compliance with these regulations does not only guarantee patient protection, but also boosts the credibility of manufacturers and enables them to access international markets [25].

Quality control in the manufacture of medical devices

Medical device manufacturing quality control (QC) is an essential part of the general quality management system that helps to make sure that the products always correspond to predetermined



specifications, safety, and regulatory requirements [26]. Since medical devices directly affect the health of patients, any small flaws or discrepancies may have significant consequences. Quality control is thus not only a last step in the manufacturing process but a systematic process that is carried out in the entire manufacturing lifecycle. In simple terms, quality control is a process of using standardized processes to check and ensure the quality of raw materials, components and finished products. This starts with checking and testing of materials coming in to make sure that they are up to specifications [27]. Suppliers are usually assessed and passed on stringent requirements and materials are physically, chemically or biologically tested and then approved to be used in production.

This is the first step to ensure defects do not creep into the production process at an early stage. In-process quality control is applied during the production process to track significant parameters and identify deviation in real-time. Such measures can be dimensional checks, environmental checks, and component testing [28]. Methods like statistical process control (SPC) are common in the analysis of process data and establishing trends that could be a sign of possible quality problems. Early detection of these problems will enable manufacturers to decrease waste, enhance efficiency, and ensure product consistency [29].

Another crucial part of quality control is the final product testing. Prior to releasing a medical device into the market, it has to be subjected to intensive testing to ensure that it is able to perform, safe and in line with the regulations. This can involve mechanical test, electrical safety test, biocompatibility test and sterility test depending on the type of the device. In case of sterile products, sterilization is done under validated procedures and routine monitoring is observed to ascertain its effectiveness [30]. Only those products that satisfy all the acceptance requirements are accepted to be distributed. Quality control in the production of medical devices is strongly associated with risk management. The manufacturers should be able to know the possible risks involved in product defect and take control measures to help in reducing the risk [31]. The evaluation of the possible points of failures and the ranking of corrective actions are often conducted with the help of tools like failure mode and effects analysis (FMEA).



This is proactive to avoid defects instead of merely identifying them once they have been experienced. Documentation and traceability are also essential elements of quality control. All manufacturing and testing processes should be documented, and they should have a clear audit trail to regulatory authorities and in-house reviews [32]. The traceability systems enable manufacturers to trace the origin of each product to the raw materials and the production environment and can easily investigate and recall in case of any need. Quality control has been improved due to the development of automation and digital technologies. Machine vision, automated inspection systems and real time data analytics have enhanced accuracy, decreased human error and enhanced efficiency [33]. Such technologies assist in a more proactive and data-driven quality management approach.

Process Validation in Medical Device Manufacturing

Process validation is an essential aspect of the manufacturing of medical devices that makes sure that the production processes produce products that have the required quality. As opposed to quality control which concentrates on detection of defects in the completed products, process validation concentrates on the integration of quality into the process. This proactive method is particularly significant in the medical equipment sector where some operations, e.g., sterilization, welding, or coating, cannot be completely tested by the end product itself [34].

Process validation is normally conducted in three key phases namely, Installation Qualification (IQ), Operational Qualification (OQ) and Performance Qualification (PQ). Installation Qualification- This is a process that ensures that equipment and systems are installed correctly as per the manufacturer requirements and environmental standards. Operational Qualification aims at verifying the equipment against the specific operating conditions to make sure that it works as intended [35]. Performance Qualification ensures that the process is capable of constantly delivering products that satisfy all the quality requirements when the process is under normal production conditions. Collectively, these steps offer an excellent degree of confidence that the production procedure is trustworthy and repeatable [36].



Process Validation in Medical Device Manufacturing




Figure 2. Process Validation in Medical Device Manufacturing

One of the major features of process validation is the identification and control of critical process parameters (CPPs) and critical quality attributes (CQAs). CPPs are the variables which may influence the result of the manufacturing process, e.g., temperature, pressure, or time, whereas CQAs are those qualities which are measurable and determine the quality of the products. Knowing the correlation between CPPs and CQAs, manufacturers may come up with agreeable ranges and introduce controls to achieve consistency. Documentation is an important part of process validation [37]. Comprehensive validation procedures and documentation have to be drawn up and kept to show adherence to regulatory standards. These documents contain validation plans, test results, deviations and conclusions which give a full report on audits and inspection [38]. This evidence is necessary to regulatory authorities to make sure that validated processes are scientifically sound and consistently implemented.

Process validation cannot be a onetime activity but a lifecycle approach. In the case of routine production, continued verification of the process is done to verify that the verified state is held



over time. Any alterations in equipment, materials or procedures should be well considered and possibly re-validated to ensure that the quality of the product suffers no harm [39]. Process validation plays a crucial role in ensuring a consistent quality of the products, reducing risks, and achieving regulatory compliance in manufacturing medical devices. Manufacturers can also be more confident in creating safe and effective devices by considering process reliability and control [40].

Interrelationship between Process Validation and Quality Control

The two key inextricable parts of quality assurance in the production of medical devices are quality control and process validation. Although they are used differently, the combination of their use will guarantee that products are not only always produced but also adhere to high safety and other standards of operations. Their relationship needs to be understood in order to come up with an effective and robust quality management system. Quality control is concerned more with detection of defects by inspection, testing and verifying processes [41]. It is a reactive system that detects nonconformities in raw materials, in-process materials, or finished products. Quality control, through different testing procedures, including visual inspection, functional testing and laboratory testing, only products that pass the specified requirements are issued to the market [42]. Nevertheless, quality control alone can be unproductive since it might result in more wastage, re-work, and delays in production in case defects are found during the later stages of the process.

Process validation on the other hand is a proactive approach that seeks to minimize defects by making certain that manufacturing processes are in a position to produce goods of the right quality as desired on a consistent basis. It helps to provide recorded facts that a process when performed under specific parameters will give consistent and repeatable results [43]. Process validation by determining the key process parameters and regulating them effectively will reduce variability and lessen the risk of defects being produced in the first place.

These two concepts are related in that they complement each other. Process validation provides the basis of an unchanging and regulated manufacturing environment, and quality control provides a monitoring system that ensures that the process is still functioning according to specification.



That is, validation and quality control give confidence in the process and continuous assurance of the quality of the product, respectively [44]. They form a feedback loop together that facilitates continuous improvement. As an exemplification, when the quality control testing indicates repetition of defects in a product, this can be taken to show that there is lack of control over the manufacturing process or some parameters are beyond acceptable limits [45]. In this instance, the results of quality control may initiate the re-examination or re-testing of the procedure. Likewise, a validated process is able to ease the workload of quality control by reducing the rate and severity of defects, resulting in a more efficient manufacturing process and reduced costs.

Documentation, data analysis and regulatory compliance are also highly interdependent in both quality control and process validation. They need to have comprehensive documentation of procedures, test results, deviations and corrective measures, which are crucial in audits and inspections. In addition, they are both also part of the risk management approaches, as they assist manufacturers to detect, evaluate, and address possible threats to the quality of the product and patient safety [46]. Quality control and process validation are dependent rather than independent components of a comprehensive quality system. Their combination makes sure that medical devices are produced consistently, to meet regulatory requirements, and finally to deliver safe and effective services to the patient [47].

Tools and Technologies for Quality Assurance

The development of tools and technologies has greatly reinforced quality assurance practices in the production of medical devices. Contemporary manufacturing settings are based not just on the traditional inspection procedures, but also rely on advanced digital tools, automated tools, and data-driven strategies to guarantee the steady quality of the product, compliance with the regulations, and efficiency of processes [48]. These technologies aid in the quality control as well as process validation through accuracy, elimination of human error and help to monitor the production processes in real-time.

Statistical Process Control (SPC) is one of the most common quality assurance tools. SPC refers to the statistical technique to control and monitor manufacturing processes. The ability of



manufacturers to analyze process data using control charts and trend analysis enables detection of variations early and corrective measures before defects can be introduced. This is proactive and can be used to stabilize processes at hand and make sure that production is within set limits [49]. Modern manufacturing of medical devices is also very dependent on the use of automation and robotics. The assembly, inspection, packaging and labeling are processed with automated systems, decreasing the use of manual labor and the possibility of human error [50]. Another example is that with automated vision inspection systems, it is possible to identify the microscopic defects in components which the human eye may not perceive.




Figure 3. Impact of Quality Assurance Tools in Medical Device Manufacturing

Robotics provides predictability in monotonous work, particularly in manufacturing processes that are very sensitive like the production of implantable devices and diagnostic equipment. The practice of quality assurance has been further changed with the digital monitoring systems and Industrial Internet of Things (IIoT) technologies. These systems allow constant checking of critical process parameters like temperature, pressure, humidity and machine performance [51]. The real-



time data collection will enable the manufacturers to react instantly to the deviation and minimize downtime as well as defective production batches. Also, centralized data storage and analysis are made possible by cloud-based platforms to enhance traceability and decision-making [52].

Modern quality assurance systems must also include data integrity and traceability. Manufacturers should make sure that all the production and testing data is correct, complete and safe. eQMS are widely applied in managing documentation, change management, corrective and preventive action (CAPA) and audit trails. Such systems will help to observe the regulatory requirements and make the process of audits and inspections by the authorities efficient [53].

Machine learning and artificial intelligence are some of the advanced technologies that are being introduced in the medical device industry. These technologies can process large amounts of data to determine details of latent patterns, anticipate possible failures, and streamline manufacturing processes [54]. An example of this is predictive maintenance that aids in minimizing the downtime of equipment by determining when machines are likely to become nonfunctional. Quality assurance tools and technologies have long since been developed out of manual inspection systems, moving to highly automated and data-driven tools and technologies. These innovations help improve the reliability, efficiency and accuracy of the process of manufacturing medical devices and eventually lead to increased quality of products and better patient safety [55].

Quality Control and Process Validation problems

Safety, efficacy and compliance with the regulations of the medical devices can be guaranteed through quality control and process validation, yet the implementation and maintenance of these systems cause a number of important challenges. These issues are due to the growing complexity of products, stringent regulatory demands, technological changes, and the necessity to have the global standardization of manufacturing processes [56]. Variability of materials, processes, and human involvement is one of the main issues of quality control. Product consistency can be influenced even by slight changes in raw materials or environmental conditions. In medicine where the accuracy of devices used is vital, even minute modifications can cause a malfunction or failure



in the device [57]. The ability of such considerable production batches to maintain a consistent quality thus necessitates very controlled environments and advanced monitoring systems.

The other significant obstacle is the growing complexity of medical devices themselves. The current devices tend to combine mechanical, electrical, and software systems and thus quality evaluation is more complex. Such multi-functional devices might need to be tested with the use of advanced testing methods and interdisciplinary knowledge which may not be achieved using traditional inspection methods. Process validation is also a complex issue, as a number of interdependent processes need to be concurrently validated to guarantee system reliability [58].

Compliance with regulations is also a major challenge. Various manufacturers of medical devices are required to adhere to numerous international standards and regulatory frameworks including those developed by various countries and agencies. Such regulations are also constantly revised, and the companies have to keep on changing their quality systems [59]. Compliance is associated with a lot of documentation, auditing and validation, which may be time-consuming and resource intensive. Process validation itself is a challenge of cost, time and scalability. Validation activities are also complex to plan, test, and document which in turn may make a substantial escalation of production schedules and cost of operation. Also, new equipment, materials or any other change in process parameters might need re-validation, which is one more burden to manufacturers. This has the potential to slug product development and innovation [60].

In modern manufacturing environments, data management and integrity have also been a pressing issue. As more and more digital systems and automated data collection is used, it is necessary to ensure the accuracy, security, and traceability of the data. Any breach on data integrity may result in non-compliance with regulations and possible product recall. Knowledge and human resource development will continue to be a challenge. To be able to make quality control and validate process, it is necessary to have highly qualified personnel, who will understand engineering, regulatory requirements, and statistical analysis [61]. The number of trained professionals in this field can however be inadequate to restrict the efficiency and effectiveness of quality systems.



Future of the Medical device manufacturing

The medical devices manufacturing sector is changing at a very fast rate due to the improved technology, the rising expectations of regulations and the rising pressure of custom-made healthcare services. The future prospects in this area are likely to make a great contribution to the quality control, process validation, and general production efficiency, as well as patient safety and product performance [62]. One of the crucial emerging trends is the additional use of digital manufacturing and Industry 4.0 technologies. With sensors, real-time data analysis, and interconnected structures, the smart factories enable the monitoring of the production processes on a continual basis [63].

This will allow manufacturers to identify deviations in real-time and make data-driven decisions. There is also an increasing significance of digital twins, virtual representations of manufacturing processes or devices, which enables simulations and optimization of the processes prior to the real production process and minimizes risks and increases the accuracy of validation. The future of quality assurance is also being influenced by artificial intelligence (AI) and machine learning (ML) [64]. The technologies will be able to process extensive amounts of production data to discover latent trends, anticipate equipment breakdowns and streamline process variables. Predictive analytics can assist manufacturers to shift to the proactive quality management in place of the reactive quality control that minimizes defects and enhances the process stability [65]. The AI-based inspection systems are also enhancing accuracy in identifying microscopic features of defects that might be overlooked by conventional methods.

One more important trend is the increase of medical device personalization and patient-specific devices. Additive manufacturing (3D printing) technologies can be used to create custom implants, prostheses, and surgical instruments based on the specific anatomy of a patient. Though this enhances better treatment results, it also poses new problems in process validation and quality control, whereby every product might need different validation techniques rather than the mass production methods [66]. The issues of cybersecurity and data integrity are getting more and more relevant because manufacturing systems are getting more digital



Another emerging theme in the production of medical devices is sustainability. Businesses are looking into the use of environmentally friendly materials, processes that are efficient in energy consumption, and minimizing waste. This transformation does not only focus on environmental issues, but also in line with the global regulatory and corporate responsibility norms [67]. There is also a changing regulatory framework to keep up with the changes in technology. The digital transformation, automation, personalization, and enhanced integration of regulations will define the future of medical device manufacturing. Such tendencies will improve the quality control and process validation effectiveness and, eventually, will result in more efficient, safer, and innovative medical devices [68].

Conclusion

The pillars of quality control and process validation form the basis of medical device manufacturing, where products are always safe, perform well, and comply with the regulatory requirements. Throughout the manufacturing lifecycle, these two interconnected systems work together to minimize risks, enhance product reliability, and safeguard patient health. The need to ensure high quality systems cannot be overemphasized in an industry where even slight defects could have severe clinical effects.

Quality control is the main tool of identifying issues and making sure every product is of a specified quality and conforms to the established standards. It provides real time feedback on product quality at different production stages through systematic inspection, testing and monitoring. This involves the evaluation of raw materials, in-process inspection and testing of the final products. Nonetheless, though QC is crucial in detecting nonconformities, it is essentially a reactive tool. It discovers issues once they have happened and this may result in wastage, slowing down production, and a higher cost of production unless it is accompanied by preventive strategies.

Process validation is an extension of quality control in that it offers a proactive model that guarantees that manufacturing processes can be used to continuously manufacture high quality devices. Validation minimizes variability and decreases the chance of defects by providing documented evidence of process stability and reproducibility that reduces variability and



minimizes the probability of defects arising in the first place. The systematic method of Installation Qualification (IQ), Operational Qualification (OQ), and Performance Qualification (PQ) makes sure that equipment, processes, and systems can operate reliably under specified conditions. The combination of these validation steps creates a scientific foundation of reproducible manufacturing results.

The combination of quality control and process validation result in the formation of a complete quality management system, which contributes to continual improvement. Depending on the recurrence of problems identified by QC, it may be an indicator that there was a need to review or re-validated the process. On the same note, validated processes minimize the workload on QC since the number of defects is minimized, and the results are more predictable. This process feedback loop is critical in ensuring long-term process stability, and efficiency.

The significance of these systems is also supported by regulatory compliance. International standards like ISO requirements, FDA regulations and EU Medical Device Regulations require stringent compliance to quality management and validation practices. Compliance does not only guarantee access to the market legally, but also fosters trust between healthcare providers and patients. Manufacturers should thus keep a lot of documentation, traceability and audit preparedness in order to be able to show conformity at any given time. Quality assurance of medical device manufacturing is also making a future with technology. By automation, artificial intelligence and real-time data analytics, more accurate monitoring and predictive quality management is now possible. Both QC and process validation are improved with these innovations, with the aim of increasing accuracy, reducing human error, and providing proactive decision-making.

Regardless of these developments, there are still some challenges such as the growing complexity of products, the changing regulatory demands, and highly skilled professionals are required. To overcome these issues, it is necessary to constantly invest in training, technology, and streamlining of the process. Quality control and process validation are not autonomous functions, but rather the combined elements of the single quality strategy. Their joint use guarantees manufacture of medical devices in a consistent, safe, and efficient way. With the industry still in the transition



phase, reinforcement of these systems will still be necessary in the pursuit of regulatory compliance, better patient outcomes, and innovation in medical technology.

References

- [1]. Harer J. Process validation for medical device manufacturing. In *Medical Devices and In Vitro Diagnostics: Requirements in Europe 2023* Jun 16 (pp. 1-30). Cham: Springer International Publishing.
- [2]. Kumar R, Kumar S. ROLE OF QUALIFICATION AND VALIDATION IN MEDICAL DEVICE INDUSTRY: A STUDY. *I-Manager's Journal on Future Engineering & Technology*. 2021 Nov 1; 17(2).
- [3]. McDermott O, Conroy N, Thenarasu M, Duarte S. Digitalised validation systems as an enabler for quality 4.0 within a medical device manufacturer. *Sustainable Futures*. 2024 Dec 1; 8:100383.
- [4]. Lange R, Schnor T. Product quality, quality control and validation. In *Practical pharmaceuticals: an international guideline for the preparation, care and use of medicinal products 2023* Jun 16 (pp. 767-783). Cham: Springer International Publishing.
- [5]. Ermer J, Nethercote PW, editors. *Method validation in pharmaceutical analysis: A guide to best practice*. John Wiley & Sons; 2025 Mar 6.
- [6]. Wang H, Chen Y, Wang L, Liu Q, and Yang S, Wang C. Advancing herbal medicine: enhancing product quality and safety through robust quality control practices. *Frontiers in pharmacology*. 2023 Sep 25; 14:1265178.
- [7]. Viceconti M, Pappalardo F, Rodriguez B, Horner M, Bischoff J, Tshinanu FM. In silico trials: Verification, validation and uncertainty quantification of predictive models used in the regulatory evaluation of biomedical products. *Methods*. 2021 Jan 1; 185:120-7.
- [8]. Goldsack JC, Coravos A, Bakker JP, Bent B, Dowling AV, Fitzer-Attas C, Godfrey A, Godino JG, Gujar N, Izmailova E, Manta C. Verification, analytical validation, and clinical validation (V3): the foundation of determining fit-for-purpose for Biometric Monitoring Technologies (BioMeTs). *npj digital Medicine*. 2020 Apr 14;3(1):55.



- [9]. Arden NS, Fisher AC, Tyner K, Yu LX, Lee SL, Kopcha M. Industry 4.0 for pharmaceutical manufacturing: Preparing for the smart factories of the future. *International journal of pharmaceutics*. 2021 Jun 1;602:120554.
- [10]. World Health Organization. Quality assurance of pharmaceuticals: a compendium of guidelines and related materials, Volume 1. Good practices and related regulatory guidance. World Health Organization; 2024 Oct 24.
- [11]. Senoner J, Netland T, Feuerriegel S. Using explainable artificial intelligence to improve process quality: evidence from semiconductor manufacturing. *Management Science*. 2022 Aug;68(8):5704-23.
- [12]. Almeida J, Bezerra M, Markl D, Berghaus A, Borman P, Schlindwein W. Development and validation of an in-line API quantification method using AQbD principles based on UV-vis spectroscopy to monitor and optimise continuous hot melt extrusion process. *Pharmaceutics*. 2020 Feb;12(2):150.
- [13]. Bitkina OV, Kim HK, Park J. Usability and user experience of medical devices: An overview of the current state, analysis methodologies, and future challenges. *International Journal of Industrial Ergonomics*. 2020 Mar 1;76:102932.
- [14]. Balekundri A, Mannur V. Quality control of the traditional herbs and herbal products: a review. *Future Journal of Pharmaceutical Sciences*. 2020 Oct 5;6(1):67.
- [15]. Joshi G, Jain A, Araveeti SR, Adhikari S, Garg H, Bhandari M. FDA-approved artificial intelligence and machine learning (AI/ML)-enabled medical devices: an updated landscape. *Electronics*. 2024 Jan 24;13(3):498.
- [16]. MacNeill AJ, Hopf H, Khanuja A, Alizamir S, Bilec M, Eckelman MJ, Hernandez L, McGain F, Simonsen K, Thiel C, Young S. Transforming the medical device industry: road map to a circular economy: study examines a medical device industry transformation. *Health Affairs*. 2020 Dec 1;39(12):2088-97.
- [17]. VARMA N, CHANDRA R, IYER S, NARAYANAN P. Data-Driven Software Quality Assurance: Leveraging Machine Learning for Risk Prediction and Test Optimization. *International Journal of Mathematical Analysis and Research*. 2026 Jan 2;2(1):1-8.



- [18]. Pradhan B, Bhattacharyya S, Pal K. IoT-based applications in healthcare devices. *Journal of healthcare engineering*. 2021;2021(1):6632599.
- [19]. Zonnenshain A, Kenett RS. Quality 4.0—the challenging future of quality engineering. *Quality engineering*. 2020 Oct 1;32(4):614-26.
- [20]. Gerke S, Babic B, Evgeniou T, Cohen IG. The need for a system view to regulate artificial intelligence/machine learning-based software as medical device. *NPJ digital medicine*. 2020 Apr 7;3(1):53.
- [21]. Alzoubi HM, In'airat M, Ahmed G. Investigating the impact of total quality management practices and Six Sigma processes to enhance the quality and reduce the cost of quality: the case of Dubai. *International journal of business excellence*. 2022;27(1):94-109.
- [22]. Mehta M, Bui TA, Yang X, Aksoy Y, Goldys EM, Deng W. Lipid-based nanoparticles for drug/gene delivery: an overview of the production techniques and difficulties encountered in their industrial development. *ACS Materials au*. 2023 Aug 21;3(6):600-19.
- [23]. Feng J, Phillips RV, Malenica I, Bishara A, Hubbard AE, Celi LA, Pirracchio R. Clinical artificial intelligence quality improvement: towards continual monitoring and updating of AI algorithms in healthcare. *NPJ digital medicine*. 2022 May 31;5(1):66.
- [24]. Đorđević S, Gonzalez MM, Conejos-Sánchez I, Carreira B, Pozzi S, Acúrcio RC, Satchi-Fainaro R, Florindo HF, Vicent MJ. Current hurdles to the translation of nanomedicines from bench to the clinic. *Drug delivery and translational research*. 2022 Mar;12(3):500-25.
- [25]. World Health Organization. Data quality assurance. Module 1. Framework and metrics. World Health Organization; 2023 Feb 14.
- [26]. Fleming GA, Petrie JR, Bergenstal RM, Holl RW, Peters AL, Heinemann L. Diabetes digital app technology: benefits, challenges, and recommendations. A consensus report by the European Association for the Study of Diabetes (EASD) and the American Diabetes Association (ADA) Diabetes Technology Working Group. *Diabetes care*. 2020 Jan 1;43(1):250-60.
- [27]. Farid SS, Baron M, Stamatis C, Nie W, Coffman J. Benchmarking biopharmaceutical process development and manufacturing cost contributions to R&D. In *MAbs* 2020 Jan 1 (Vol. 12, No. 1, p. 1754999). Taylor & Francis.



- [28]. Sader S, Husti I, Daroczi M. A review of quality 4.0: definitions, features, technologies, applications, and challenges. *Total quality management & business excellence*. 2022 Jul 4;33(9-10):1164-82.
- [29]. Zhang J, Zhang ZM. Ethics and governance of trustworthy medical artificial intelligence. *BMC medical informatics and decision making*. 2023 Jan 13;23(1):7.
- [30]. Amann J, Blasimme A, Vayena E, Frey D, Madai VI, Precise4Q Consortium. Explainability for artificial intelligence in healthcare: a multidisciplinary perspective. *BMC medical informatics and decision making*. 2020 Nov 30;20(1):310.
- [31]. Farouk A, Alahmadi A, Ghose S, Mashatan A. Blockchain platform for industrial healthcare: Vision and future opportunities. *Computer Communications*. 2020 Mar 15;154:223-35.
- [32]. Ali MM, Hashim N, Abd Aziz S, Lasekan O. Principles and recent advances in electronic nose for quality inspection of agricultural and food products. *Trends in Food Science & Technology*. 2020 May 1;99:1-0.
- [33]. Omoumi P, Ducarouge A, Tournier A, Harvey H, Kahn Jr CE, Louvet-de Verchère F, Pinto Dos Santos D, Kober T, Richiardi J. To buy or not to buy—evaluating commercial AI solutions in radiology (the ECLAIR guidelines). *European radiology*. 2021 Jun;31(6):3786-96.
- [34]. Salvi S. Quality assurance and quality control for project effectiveness in construction and management. *International Journal of Engineering Research & Technology (IJERT)*. 2020 Jan 1.
- [35]. Neira-Rodado D, Ortíz-Barrios M, De la Hoz-Escorcía S, Paggetti C, Noffrini L, Fratea N. Smart product design process through the implementation of a fuzzy Kano-AHP-DEMATEL-QFD approach. *Applied sciences*. 2020 Mar 5;10(5):1792.
- [36]. Raji ID, Smart A, White RN, Mitchell M, Gebru T, Hutchinson B, Smith-Loud J, Theron D, Barnes P. Closing the AI accountability gap: Defining an end-to-end framework for internal algorithmic auditing. In *Proceedings of the 2020 conference on fairness, accountability, and transparency* 2020 Jan 27 (pp. 33-44).



- [37]. Al-Saffar NG, Obeidat AM. The effect of total quality management practices on employee performance: The moderating role of knowledge sharing. *Management science letters*. 2020;10(1):77-90.
- [38]. Rowan NJ, Laffey JG. Challenges and solutions for addressing critical shortage of supply chain for personal and protective equipment (PPE) arising from Coronavirus disease (COVID19) pandemic–Case study from the Republic of Ireland. *Science of the Total Environment*. 2020 Jul 10;725:138532.
- [39]. Kumar S, Gopi T, Harikeerthana N, Gupta MK, Gaur V, Krolczyk GM, Wu C. Machine learning techniques in additive manufacturing: a state of the art review on design, processes and production control. *Journal of Intelligent Manufacturing*. 2023 Jan;34(1):21-55.
- [40]. Acevedo-De-los-Ríos A, Rondinel-Oviedo DR. Impact, added value and relevance of an accreditation process on quality assurance in architectural higher education. *Quality in Higher Education*. 2022 May 4;28(2):186-204.
- [41]. Zaszczyńska A, Gradys A, Sajkiewicz P. Progress in the applications of smart piezoelectric materials for medical devices. *Polymers*. 2020 Nov 22;12(11):2754.
- [42]. Szymczyk-Ziółkowska P, Łabowska MB, Detyna J, Michalak I, Gruber P. A review of fabrication polymer scaffolds for biomedical applications using additive manufacturing techniques. *Biocybernetics and Biomedical Engineering*. 2020 Apr 1;40(2):624-38.
- [43]. Rak R, Quinn P. Enhancing Digital Health Innovation in the EU with Effective Industrial Strategy Policies-A Focus on Wearable Medical Devices.
- [44]. Jonusauskas L, Baravykas T, Butkutė A, Andriukaitis D, Gričius H, Tiškūnas T, Andrijev D. 3D laser microfabrication of medical devices. *Transactions on Additive Manufacturing Meets Medicine*. 2020 Sep 9;2(1).
- [45]. McCarthy DJ, Hinchey EP, O'Dowd NP, McCarthy CT, McMorow D. Using model based design as an enabler for digital validation of discrete state machines in regulated manufacturing environments. *Procedia Manufacturing*. 2021 Jan 1;55:365-70.
- [46]. Valls Esteve A, Lustig Gainza P, Adell Gomez N, Tejo Otero A, Englí Rueda M, Julian Alvarez E, Navarro Sureda O, Fenollosa i Artés F, Rubio Palau J, Krauel L, Munuera J. A state-of-the-art guide about the effects of sterilization processes on 3D-printed materials



- for surgical planning and medical applications: A comparative study. *International journal of bioprinting*. 2023;9(5).
- [47]. Singh BJ, Sehgal R. Applications of 3D printing, rapid prototyping, and digital manufacturing for low carbon development. In *Sustainability in smart manufacturing 2024* May 14 (pp. 37-60). CRC Press.
- [48]. Capach NH, Zigdon N, Payne TA, Neukam JD, Choi Y, Park HJ, Shapiro WH, Svirsky MA. Cochlear Implants and the Aided Audiogram: A Retrospective study comparing performance across device Manufacturers. *Audiology Research*. 2025 Jul 2;15(4):79.
- [49]. Bhatia A, Sehgal AK. Additive manufacturing materials, methods and applications: A review. *Materials Today: Proceedings*. 2023 Jan 1;81:1060-7.
- [50]. Moon KJ, Dong CW, Ham DH, Park WT. Ultra-Miniature MEMS Sensor Packaging Process for Intravascular Medical Device. *International Journal of Precision Engineering and Manufacturing*. 2025 Oct;26(10):2739-48.
- [51]. Qiu H, Si Z, Luo Y, Feng P, Wu X, Hou W, Zhu Y, Chan-Park MB, Xu L, Huang D. The mechanisms and the applications of antibacterial polymers in surface modification on medical devices. *Frontiers in bioengineering and biotechnology*. 2020 Nov 11;8:910.
- [52]. Calderone D, Cesarelli G, Cesarelli M, Iuppariello L, Guida P, Casaburi A, Romano G, Amato F, Clemente F. Optimization of 3D fused deposition modeling printing process for the manufacturing of devices for medical use. In *2023 IEEE International Conference on Metrology for extended Reality, Artificial Intelligence and Neural Engineering (MetroXRINE)* 2023 Oct 25 (pp. 484-489). IEEE.
- [53]. Siewert S, Bohne E, Grossmann S, Schmidt W, Grabow N, Stahnke T, Guthoff R, Stiehm M, Schmitz KP. Femtosecond laser manufacturing technology for microstents as innovative medical devices to treat open-angle glaucoma. *Current Directions in Biomedical Engineering*. 2022 Aug 1;8(2).
- [54]. Gao Z, Zhou Y, Zhang J, Foroughi J, Peng S, Baughman RH, Wang ZL, Wang CH. Advanced energy harvesters and energy storage for powering wearable and implantable medical devices. *Advanced Materials*. 2024 Oct;36(42):2404492.



- [55]. Zhang C, Li Y, Kang W, Liu X, Wang Q. Current advances and future perspectives of additive manufacturing for functional polymeric materials and devices. *SusMat*. 2021 Mar;1(1):127-47.
- [56]. Shineh G, Mobaraki M, Perves Bappy MJ, Mills DK. Biofilm formation, and related impacts on healthcare, food processing and packaging, industrial manufacturing, marine industries, and sanitation—a review. *Applied Microbiology*. 2023 Jun 26;3(3):629-65.
- [57]. De Jong WH, Carraway JW, Geertsma RE. In vivo and in vitro testing for the biological safety evaluation of biomaterials and medical devices. In *Biocompatibility and performance of medical devices 2020* Jan 1 (pp. 123-166). Woodhead Publishing.
- [58]. Guzzi EA, Tibbitt MW. Additive manufacturing of precision biomaterials. *Advanced materials*. 2020 Apr;32(13):1901994.
- [59]. Wang Y, Cui H, Esworthy T, Mei D, Wang Y, Zhang LG. Emerging 4D printing strategies for next-generation tissue regeneration and medical devices. *Advanced Materials*. 2022 May;34(20):2109198.
- [60]. Deng X, Gould M, Ali MA. A review of current advancements for wound healing: Biomaterial applications and medical devices. *Journal of Biomedical Materials Research Part B: Applied Biomaterials*. 2022 Nov;110(11):2542-73.
- [61]. Kasoju N, Remya NS, Sasi R, Sujesh S, Soman B, Kesavadas C, Muraleedharan CV, Varma PH, Behari S. Digital health: trends, opportunities and challenges in medical devices, pharma and bio-technology. *CSI Transactions on ICT*. 2023 Apr;11(1):11-30.
- [62]. Muehlematter UJ, Daniore P, Vokinger KN. Approval of artificial intelligence and machine learning-based medical devices in the USA and Europe (2015–20): a comparative analysis. *The Lancet Digital Health*. 2021 Mar 1;3(3):e195-203.
- [63]. Ghomi ER, Khosravi F, Neisiany RE, Singh S, Ramakrishna S. Future of additive manufacturing in healthcare. *Current Opinion in Biomedical Engineering*. 2021 Mar 1;17:100255.
- [64]. Lepasepp TK, Hurst W. A systematic literature review of industry 4.0 technologies within medical device manufacturing. *Future Internet*. 2021 Oct 13;13(10):264.



- [65]. Aronson JK, Heneghan C, Ferner RE. Medical devices: definition, classification, and regulatory implications. *Drug safety*. 2020 Feb 1;43(2):83-93.
- [66]. Beckers R, Kwade Z, Zanca F. The EU medical device regulation: Implications for artificial intelligence-based medical device software in medical physics. *Physica Medica*. 2021 Mar 1;83:1-8.
- [67]. Li C, Pisignano D, Zhao Y, Xue J. Advances in medical applications of additive manufacturing. *Engineering*. 2020 Nov 1;6(11):1222-31.
- [68]. Festas AJ, Ramos A, Davim JP. Medical devices biomaterials—A review. *Proceedings of the institution of mechanical engineers, part L: Journal of Materials: Design and Applications*. 2020 Jan;234(1):218-28.